

All data taken at Pacific Northwest National Laboratory (PNNL)
Operator: Steven W. Sharpe and Robert L. Sams, sw.sharpe@pnl.gov
Version 1.0, February, 01

Composite spectrum for CS₂_5T

Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K

Equivalent concentration x path-length of composite spectrum: 3.135×10^{-6} grams/liter-meter

Sample Conditions-

- Chemical name and CAS number: Carbon disulfide, carbon bisulfide, CS₂ : [75-15-0]
- Physical properties: M.W. 76.14 amu, F.P. -111 C, B.P. 46.3 C
- Supplier and stated purity: Aldrich, 99.9+% (HPLC grade)
- Sample class: I (PNNL scale).
- Temperature of sample: 5.04 ± 0.02 C
- Diluent: Sample back filled with ultra high purity nitrogen to 760 ± 5 Torr
- Individual samples 0.27754, 5.9315, 0.65859, 12.31, 0.49008, 26.74, 1.21852 and 2.6343 Torr.
Path length = 19.94 cm. Final data is a composite spectrum.
- Preparation: Multiple freeze-thaw cycles at -80 C to remove air.

Instrument Parameters-

- Bruker-66V FTIR, temperature controlled environment, evacuated optics bench
- Modified to include second aperture, between interferometer output and sample cell. This substantially reduces both "ghosting" and warm aperture effects.
- Spectral range: 6,500 to 600 cm⁻¹ (1.534 to 16.667 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112 cm⁻¹
- Spectral intervals after FFT: 0.06 cm⁻¹
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0 cm⁻¹

Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram ($\alpha = 0.85$, $\beta = 530$)
- Composite spectrum created from 8 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by T² (transmission squared), all absorbance values > 1.6 are given zero weight
- Calculated and estimated errors: Type A = 0.34%, Type B = 3%
- Frequency correction (already applied): $V(\text{corrected}) = V(\text{instrument}) * 0.999997 + 5.18 \times 10^{-4}$
- Axis units: X=wavenumbers (cm⁻¹), Y=Absorbance (base-10)
- Trace water and carbon dioxide vapor features removed via spectral subtraction
- Baseline correction via 7th order polynomial subtraction