

All data taken at Pacific Northwest National Laboratory (PNNL)

Operators: Steven W. Sharpe, Timothy J. Johnson and Robert L. Sams : sw.sharpe@pnl.gov

Version 1.0, May, 01

Composite spectrum for DMS_50T

Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K

Equivalent concentration x path-length of composite spectrum: 2.558×10^{-6} grams/liter-meter

Sample Conditions-

- Chemical name and CAS number: Dimethyl sulfide, methyl sulfide, H_3CSCH_3 : [75-18-3]
- Physical properties: fw=62.13 g/mole, fp=-83° C, bp=37.5° C
- Supplier and stated purity: Aldrich, 99+%. Observed an unknown contamination peak (probably a ketone) at 1755 cm^{-1} .
- Sample class: I (PNNL scale).
- Temperature of sample: $50.00 \pm 0.02 \text{ C}$
- Diluent: Sample back filled with ultra high purity nitrogen to $760 \pm 5 \text{ Torr}$
- Individual samples at 1.3460, 16.10, 3.0216, 6.2576, 0.75550 and 70.40 Torr. Path length = 19.96 cm. Final data is a composite spectrum.
- Preparation: Sample placed over CaSO_4 to remove water. Multiple freeze-thaw cycles at -50 C to remove air.

Instrument Parameters-

- Bruker-66V FTIR, temperature controlled environment, evacuated optics bench
- Modified to include second aperture, between interferometer output and sample cell. This substantially reduces both “ghosting” and warm aperture effects.
- Spectral range: $6,500$ to 600 cm^{-1} (1.534 to 16.667 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112 cm^{-1}
- Spectral interval after 2X zero-filling interferogram and FFT: 0.06 cm^{-1}
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0 cm^{-1}

Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram ($\alpha = 0.90$, $\beta = 500$)
- Composite spectrum created from 6 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by T^2 (transmission squared), all absorbance values > 1.6 are given zero weight
- Calculated and estimated errors: Type A = 1.13%, Type B = 3%
- Frequency correction (already applied): $V(\text{corrected}) = V(\text{instrument}) * 0.999998 + 1.287 \times 10^{-4}$
- Axis units: X=wavenumbers (cm^{-1}), Y=Absorbance (base-10)
- Trace carbon dioxide features removed via spectral subtraction
- Baseline correction via 7th order polynomial subtraction