All data taken at Pacific Northwest National Laboratory (PNNL)
Operators: Steven W. Sharpe, Timothy J. Johnson and Robert L. Sams : sw.sharpe@pnl.gov
Version 1.0, February, 04

Composite spectrum for F13B1_50T
Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K
Equivalent concentration x path-length of composite spectrum: 6.1307×10^{-6} grams/liter-meter
Trace unidentified impurity observed at 1150 cm^{-1}.

Sample Conditions-

- Chemical name and CAS number: Bromotrifluoromethane, Freon-13B1, Halon-1301, trifluorobromomethane, $CBrF_3$: [75-63-8]
- Physical properties: fw=148.9102 g/mole, fp=–168° C, bp=-57.8° C
- Supplier and stated purity: SynQuest Labs, 99%
- Sample class: II (PNNL scale). Extremely strong absorber requiring small partial pressures of sample.
- Temperature of sample: 49.97.00 ± 0.02 C
- Diluent: Sample back filled with ultra high purity nitrogen to 760±5 Torr
- Individual samples at 0.13062, 0.21679, 0.33803, 4.8818, 2.2104, 43.95, 18.49 and 8.9607 Torr. Path length = 19.96 cm. Final data is a composite spectrum.
- Preparation: Multiple freeze-thaw cycles at 77 K to remove air. Bulk sample kept at –55 C during transfer to sample cell to avoid water vapor contamination.

Instrument Parameters-

- Bruker-66V FTIR, temperature controlled environment, evacuated optics bench
- Modified to include second aperture, between interferometer output and sample cell. This substantially reduces both "ghosting" and warm aperture effects.
- Spectral range: 6,500 to 510 cm^{-1} (1.534 to 19.608 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112 cm^{-1}
- Spectral interval after 2X zero-filling interferogram and FFT: 0.06 cm^{-1}
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0 cm^{-1}

Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram (α=0.85, ε=530)
- Composite spectrum created from 8 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by T^2 (transmission squared), all absorbance values ≥ 1.6 are given zero weight
- Calculated and estimated errors: Type A = 0.49%, Type B ≤ 5%
- Frequency correction (already applied): V(corrected) = V(instrument)* $0.99999896+8.812 \times 10^{-4}$
- Axis units: X=wavenumbers (cm^{-1}), Y=Absorbance (base-10)

- Trace carbon dioxide features removed via spectral subtraction
- Baseline correction via 6^{th} order polynomial subtraction