

All data taken at Pacific Northwest National Laboratory (PNNL)
Operators: Steven W. Sharpe, Timothy J. Johnson and Robert L. Sams : sw.sharpe@pnl.gov
Version 1.0, October, 01

Composite spectrum for MCDS_25T

Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K

Equivalent concentration x path-length of composite spectrum: 8.966×10^{-6} grams/liter-meter

This is not a pure sample and consists of a mixture of similar compounds and isomers.

Sample Conditions-

- Chemical name and CAS number: Methylchlorodisilanes, primarily mixed isomers of tritrimethyltrichlorodisilane $(\text{CH}_3)_3\text{Si-SiCl}_3$ with small amount of dimethyltetrachlorodisilanes $(\text{CH}_3)_2\text{Si-SiCl}_4$: [CAS unavailable]
- Physical properties: fw 218 g/mole, fp=not available° C, bp 155° C
- Supplier and stated purity: Aldrich, mixed isomers
- Sample class: III (PNNL scale). Mixture of compounds and isomers
- Temperature of sample: 25.01 ± 0.02 C
- Diluent: Sample back filled with ultra high purity nitrogen to 760 ± 5 Torr
- Individual samples at 1.03620, 2.1340, 3.0190, 0.75710, 1.5284, 2.5323, 0.58352 and 1.8378 Torr. Path length = 19.94 cm. Final data is a composite spectrum.
- Preparation: Multiple freeze-thaw cycles at 0 C to remove air.

Instrument Parameters-

- Bruker-66V FTIR, temperature controlled environment, evacuated optics bench
- Modified to include second aperture, between interferometer output and sample cell. This substantially reduces both “ghosting” and warm aperture effects.
- Spectral range: 6,500 to 520 cm^{-1} (1.534 to 19.231 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112 cm^{-1}
- Spectral interval after 2X zero-filling interferogram and FFT: 0.06 cm^{-1}
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0 cm^{-1}

Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram ($\tau = 0.85$, $\nu = 530$)
- Composite spectrum created from 8 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by T^2 (transmission squared), all absorbance values > 1.6 are given zero weight
- Calculated and estimated errors: Type A = 1.11%, Type B Best effort
- Frequency correction (already applied): $V(\text{corrected}) = V(\text{instrument}) * 0.999997 + 5.18 \times 10^{-4}$
- Axis units: X=wavenumbers (cm^{-1}), Y=Absorbance (base-10)