

All data taken at Pacific Northwest National Laboratory (PNNL)

Operators: Steven W. Sharpe, Timothy J. Johnson and Robert L. Sams : sw.sharpe@pnl.gov

Version 1.0, October, 01

Composite spectrum for B2H6_50T

Effective burden of composite spectrum: 1 part-per-million-meter (ppm-meter) at 296 K

Equivalent concentration x path-length of composite spectrum: 1.1391×10^{-6} grams/liter-meter

Sample Conditions-

- Chemical name and CAS number: Diborane, boroethane, boron hydride, diborane hexahydride, B₂H₆ : [19287-45-7]
- Physical properties: fw=27.667 g/mole, fp=-165° C, bp=-92.5° C
- Supplier and stated purity: Matheson, 15% in helium, semiconductor purity
- Sample class: II (PNNL scale). Extremely reactive. Strong reducer
- Temperature of sample: 50.04 ± 0.02 C
- Diluent: Sample back filled with ultra high purity nitrogen to 760±5 Torr
- Individual samples at 1.2470, 0.54370, 15.81, 2.0522, 0.81270, 7.2466, 5.0130 and 3.5733 Torr. Path length = 19.94 cm. Final data is a composite spectrum.
- Preparation: Multiple freeze-thaw cycles at 77 K to remove air and helium.

Instrument Parameters-

- Bruker-66V FTIR, temperature controlled environment, evacuated optics bench
- Modified to include second aperture, between interferometer output and sample cell. This substantially reduces both “ghosting” and warm aperture effects.
- Spectral range: 6,500 to 600 cm⁻¹ (1.534 to 16.667 microns)
- Instrumental resolution based on maximum interferometer displacement is 0.112 cm⁻¹
- Spectral interval after 2X zero-filling interferogram and FFT: 0.06 cm⁻¹
- Interferogram zero-fill: 2X
- Apodization: Boxcar
- Phase correction: Mertz
- Beam splitter: Potassium bromide (KBr)
- IR source: Carbide glowbar (22 V)
- Scanner velocity: 60KHz (HeNe crossing frequency)
- Number of interferograms averaged per single channel spectra: 256
- Detector: Mid-band HgCdTe, photoconductive, 77K operation
- Folding limits: 15798 to 0 cm⁻¹

Post Processing and Related Parameters-

- Non-linearity detector correction (Bruker proprietary) applied to interferogram ($\alpha=0.85$, $\beta=530$)
- Composite spectrum created from 8 individual absorbance (base-10) spectra via classical least squares fit: Intercept=0, slope is fitted, individual absorbance values weighted by T² (transmission squared), all absorbance values > 1.6 are given zero weight
- Calculated and estimated errors: Type A = 1.21%, Type B = 5%
- Frequency correction (already applied): $V(\text{corrected}) = V(\text{instrument}) * 0.999997 + 5.18 \times 10^{-4}$
- Axis units: X=wavenumbers (cm⁻¹), Y=Absorbance (base-10)
- Baseline correction via 7th order polynomial subtraction